OFFICE OF THE SUPERINTENDENT
DISTRICT HOSPITAL KONDAGAON.

Tel / Fax : **07786-242255**

e-mail: cskondagaon@gmail.com

100 BED MCH HOSPITAL EOI DOCUMENT
FOR
SUPPLY & INSTALLATION
OF
MEDICAL EQUIPMENTS and FURNITURE

SECTION -I

Expression of Interest

EOI Reference No. -644 Dated: 18/08/2017

Expression of Interest ARE INVITED FROM ELIGIBLE BIDDERS AS PER THE ELIGIBILITY CRITERIA FOR **SUPPLY & INSTALATION OF MEDICAL EQUIPMENTS AND FURNITURE**

1	Period of Availability of EOI Document	From 22/08/2017 (date) to 01/09/2017 date) [Downloadable from website: **www._CGSTATE.GOV.IN/KONDAGAON.GOV.IN** In case of any bid amendment and clarification, responsibility lies with the bidders to collect the same from the above mentioned website before last date of submission of EOI document and the EOI inviting authority shall have no responsibility for any delay / omission on part of the bidder.
2	Last date & time for submission of EOI	Date: 06/09/2017 Time: up to 4 pm Address of Submission of **Bid, BROWSER& CATALOGE**: The CIVIL SURGEON CUM Hospital Superintendent , District Hospital Kondagaon Pin-494226, chhattisgarh, India (Through Speed post / Registered post only
3	Date, time and place of opening of EOI	**a) Technical Bid (Cover A) opening: 07/09/2017 (date), 03 pm time** **(time) at the address mentioned above.** **b) Financial Bid (Cover B):** *The date of opening of financial bid will be intimated to the firms found successful in the technical bid evaluation.* ***(Venue is mentioned at the address mentioned above)*** ***(Bidders / authorized representative may remain present at the time of opening of bid)***



Civil Surgeon/Superintendent

SECTION -II

IMPORTANT INSTRUCTIONS TO BE NOTED CAREFULLY BY THE EOIERS

1.	Mode of Procurement	**Through National Competitive Bidding Process. The O/o of** Civil Surgeon cum Superintendent District Hospital-kondagaon, Chhattisgarh **shall invite EOI & evaluate the same. After finalization/approval of the supplier & the rate, the purchase order shall be placed by the** Civil Surgeon cum Superintendent District Hospital-kondagaon, Chhattisgarh**.**
2.	Purchaser	Civil Surgeon cum Superintendent District Hospital-kondagaon, Chhattisgarh
3.	Consignee	Department of MCH 100 Bed Hospital kondagaon, Chhattisgarh
4.	Delivery Period	Within30 **days** from issue of the purchase order.
5.	Mode of Delivery	By Air / Road / Rail
6.	Guarantee / Warranty /CMC	**Comprehensive warranty** including all spares, maintenance etc. for a period 10***(ten) years*** from the date of installation & commissioning and ***7(seven) years***
7.	EOI Document Cost	**Rs.200000** (Detail para 6)The EOI document cost is to be submitted in the shape **Nationalized/Scheduled Bank payable at** Kondagaon or deposit of Rs. **200000** /- in Accounts Section of this hospital on all working days between the scheduled period.

08	Pre-qualification **(Eligibility Criteria)**	**A.** Manufacturing units / Importers are eligible to participate in the EOI provided, they have (i) Import License (In case of Importer only) (ii) Valid ISO certificate. (iii) Product must be ISI /CE / US FDA/IEC etc certified if specified in Technical Specification (Section V) (iv) EOIer (Manufacturer/Importer) should have proof of supply of **equipment(s) mentioned in the schedule of requirement** (executed directly by manufacturer or through distributor) of the equipment(s) /similar equipments mentioned in the schedule of requirement to any Govt. organization / Corporate Hospitals / PSU Hospitals / UN Agencies and purchase order copies in support of that in last 3years as per format at Annexure VII (Item wise) (v) Proof of annual average turnover (Manufacturers/Importer) of in the last three (3) financial years certified by the Chartered Accountant as per the format at **supported by audited balance sheet/Annual Report.** **B.** Authorized distributors on behalf of the manufacturer are eligible to participate in the EOI provided: (i) They should have proof of annual average turnover of the last three (3) financial years certified by the Chartered Accountant as per the format at **Annexure supported by audited balance sheet/Annual Report.** (ii) In addition to this, the distributor shall also submit the average annual turnover of the **manufacturer/importer** of the **item (s)** as mentioned in A (v) above. (iii) They should submit **manufacturer's authorization** to transact business on behalf of the manufacturer as per the format (iv) Proof of supply of the equipments in the schedule of requirement to any Govt. organization /Corporate Hospitals / PSU Hospitals / UN Agencies and purchase order copies in support of that in last 3years as per format at Annexure VII (Item wise) (v) The authorized distributor will submit the following documents in support of the manufacturer along with the EOI : (vi) Valid ISO certificate Valid ISI / CE / US FDA / IEC certificates of the manufacturer is specified in technical specification (Section V) **D.** The Manufacturer or the EOIer if blacklisted either by the EOI inviting authority or by any state Govt. or Central Govt. organization for the quoted item is not eligible to participate in the EOI during the period of blacklisting.



SECTION -III

TERMS AND CONDITIONS FOR SUPPLY & INSTALLATION OF MEDICAL EQUIPMENTS FOR DIFFERENT DEPARTMENT

1.1 Sealed EOIs will be received till .06/09/2017 (date) upto 4pm (time) by the office of the Civil Surgeon Cum Superintendent District Hospital–kondagaon. Any EOI received after the due date & time will be rejected / returned to the sender unopened. **The EOIs will be received through Regd. Post / / Speed Post only.** The prospective bidders may attend and clarify any doubts on the terms and conditions of the bid document.

1.2 The bidder(s) are to submit their EOIs in **separate** sealed covered envelops for **technical bid** and **commercial bid** by super scribing **Cover "A" (Technical Bid)** & **Cover "B" (Price Bid)** and both the sealed covers should be put into a **third outer Cover**, which should be super scribed as **"EOI for Supply & Installation of Medical Instruments &Equipments**

1.3 The Sealed EOIs "Cover A" (Technical Bid) submitted by the EOIers will be opened at the office of the Civil Surgeon Cum Superintendent District Hospital–kondagaon on 07/09/2017.(date) at _3 pm (time).

The EOIer or their duly authorized representatives are allowed to be present during the opening of the EOIs if they so like.

ELIGIBILITY CRITERIA

2.1 **Manufacturing units / Importers** are eligible to participate in the EOI provided, they fulfill the following conditions:

(i) Import License (In case of Importer only). In case of importers, they have to furnish the authorization from the manufacturer.

(ii) Valid ISO certificate (of the Manufacturer)

(iii) Product must be ISI/BIS /CE / US FDA etc. (valid ISI/BIS /CE /US FDA certificate) certified (As per **Section** - technical specification**)**.

(iv) EOIer (Manufacturer/Importer) should have proof of supply of equipments mentioned in the schedule of requirement (executed directly by manufacturer or through distributor) of the equipment(s) /similar equipments mentioned in the schedule of requirement to any Govt. organization / Corporate Hospitals / PSU Hospitals / UN Agencies and purchase order copies in support of that in last 3years as per format at Annexure VII (Item wise)

(v) Proof of annual average turnover (Manufacturers/Importer) of in the last three (3)financial years certified by the Chartered Accountant as per the format

(vi) Proof of compliance with IEC Certificate - technical specification) - Medical Electrical Equipments: Particular requirement for Electrical Safety of the equipments.

(vii) Manufacturing unit who has been blacklisted either by the EOI inviting authority or by any state Govt. or Central Govt. organization is not eligible to participate in the EOI for that item during the period of blacklisting. Copies of stay order(s) if any against the blacklisting should be furnished along with the bid.

2.2 **Authorized distributors** are eligible to participate in the EOI provided:

(i) They submit manufacturer's authorization from original equipment manufacturer (OEM) as per the format

(ii) They should have proof of annual average turnover of in the last three (3)financial years certified by the Chartered Accountant as per the format

(iii) In addition to this, the distributor shall also submit the average annual turnover of the **manufacturer/importer** of the **item (s)** as mentioned in 2.1 (v) & (vi) above.

(iv) Proof of supply of **the required equipment** (executed directly by manufacturer or through distributor) of the equipment(s) /similar equipments mentioned in the schedule of requirement to any Govt. organization /Corporate Hospitals / PSU Hospitals / UN Agencies and purchase order copies in support of that in last 3years as per format at Annexure VII (Item wise).

(v) The authorized distributor will submit the following documents in support of the manufacturer along with the EOI:

a) Valid ISO certificate

b) CE / US FDA / IEC certificates of the manufacturer as per technical specification.

2.3 The EOIer have to submit the EMD(s) as mentioned **in Clause 8 of Section -II** & the EOI document cost.

DOCUMENTS TO BE SUBMITTED

The following documents should be enclosed in Cover "A" (Technical Bid) by the EOIer.

All the photocopies are to be attested by a Notary Public / Gazetted Officer

TECHNICAL BID :

3.1 Checklist with detail of the documents enclosed in **Cover "A"** with **page number**. The documents should be ***serially arranged*** as per this and should be securely tied and bound.

3.2 List of Item (s) Quoted with name of the Make & Model of the item (s)

3.3 EOI document fee of Rs200000/- in shape of Demand Draft .

3.4 Earnest Money Deposit(s) as mentioned in the **Clause 8 of Section -II** in shape of Demand Draft). Details of EMD and the name of the equipment quoted should be clearly mentioned.



3.5 Details name, address, telephone no., Fax, e-mail of the manufacturer / authorized distributor / service centre / contract person / office in Odisha

3.6 The declaration form in duly signed by the EOIer before Notary Public / Executive Magistrate.

3.7 Manufacturer's Authorization Format in (In case the bidder is not the manufacturer). Importers are also required to furnish the authorization from the manufacturer.

3.8 Certificate duly filled by the Auditor / Chartered Accountant that the annual average turnover of the firm in the last 3 financial years (In case of bidders who are authorized distributors of the manufacturer). In case of authorized distributor, they will also have to submit the average annual turnover the manufacturer/importer of the item(s).

3.9 Performance Statement **(Item wise)** during the last three years towards proof of supply of the equipment(s) /similar equipments mentioned in the schedule of requirement to any Govt. organization / Corporate Hospitals / PSU Hospitals / UN Agencies. The copy of Purchase orders and certificate from the user should be furnished in support of the information provided in the performance statement.

3.10 Deviation/No Deviation Statement from Technical Specification & details of technical specification of the product **(Annexure-VIIIA & B)**

3.11 **Leaflet/Technical Brochures of the product/item offered**.

3.12 Copy of Import License by the Importer (in case of Importer).

3.13 Copy of Valid ISO certificate.

3.14 Copy of Valid ISI / CE /US FDA certificate (as per Section VI - Technical Specification).

3.15 Copy of Certificate in support of IEC certificate (Technical Specification).

3.16 TAX CLEARENCE CERTIFICATE .

3.17 The Original EOI Booklet with Conditions and the schedules signed by the EOIer at the bottom of each page with his official seal duly affixed.



3.18 Certificate in support of service center in Chhattisgarh or undertaking to set up service center in Chhattisgarh within one month from the date of installation if approved (for those who have no service centers in Chhattisgarh).

N.B: Valid means the certificate should be valid on or beyond the date of opening of EOI (Cover-A).

COVER – B (PRICE BID)

4. The price to be quoted for medical equipments should be sent in the prescribed price format in a separate sealed cover hereafter called **Cover "B" (Price Bid).**

Cover –B (Price Bid) of the EOIers who qualify in it's Technical Bid (Cover – A) and complies to EOI specification & find to be as per technical specification in Product in demonstration (if required) **will only be opened .**

4.1 The EOI format (Price Schedule) in duplicate in the prescribed form (as per **Annexure – IX**), must be submitted in Cover-B. The price of the item should be quoted inclusive of excise duty, insurance, packing, forwarding, freight (door delivery) and warranty for 3 years.

(accessories if any for installation including charges for installation/commissioning), sales GST/TAX and entry Tax charges (if any) should be quoted in a separate column. The rate should be quoted for ***each item*** both in figures and words. **In case of difference in words and figures, words will be taken into consideration for evaluation.**

4.2 The Cover "B" of EOIers who qualifies in their technical bid, will only be opened at the office of the **Civil Surgeon Cum Superintendent District Hospital– kondagaon Chhattisgarh** at a date & time which will be intimated to them.

REJECTION OF EOI

5. The EOI submitted by the bidder will be rejected, if any of the following documents are wanting / not submitted with the EOI:

(i) Import License (In case of Importer)

(ii) Manufacturer's authorization in case of distributor/importer

(iii) Earnest Money Deposit (EMD)



(iv) Annual Average Turnover of Rs.2.5 Crore or more in the last 3 financial years as per Annexure –VI. Valid ISO certificate of Manufacturer

(v) Valid ISI / CE / US FDA certificate of the manufacturer as per Section VI – Technical Specification.

(vi) IEC Certificate of the manufacturer as per as per Section VI – Technical Specification.

(vii) Proof of supply/ installation (executed directly by manufacturer or through distributor) of the equipment(s) / similar equipments mentioned in the schedule of requirement to any Govt. Organization / Corporate Hospitals / PSU Hospitals / UN Agencies and certificate in support of that from the user during the **last three years** (Item wise)

(viii) Major deviations from the technical specification of the item(s) as per EOI.

(ix) Price bid / quoted rate with signature and seal (SOFT and Hard Copy).

EARNEST MONEY DEPOSIT

6.1 The amount of Earnest Money Deposit required is mentioned in the Section-II. The Earnest Money Deposit will be submitted in the shape of **demand Draft only** in favour of **" Jeevan Deep samity District Hospital–kondagaon "** Chhattisgarh ,from any Nationalized/Scheduled Bank payable at **Kondagaon**

6.2 The EMD of the unsuccessful EOIers will be returned back without interest after placement of purchase order to the successful EOIer and EMD of successful EOIer will be returned after submission of performance security (ies)

6.3 The EMD will be forfeited if the EOIer withdraws its EOI / furnish forged documents which is found during bid evaluation OR doesn't sign the contract / doesn't furnish performance security / doesn't supply the items (in case of successful bidder) within the stipulated time period.

PERFORMANCE SECURITY & AGREEMENT

7.1 The agreement (**as per Annexure**) will be signed between the supplier, manufacturer and the purchaser and will be kept by the purchaser.



7.2 The performance Security Money will be returned back to the EOIer without interest after the expiry of the warranty period i.e. two years after the date of installation & signing of the agreement.

7.3 Security money will be forfeited if there is any violation of the EOI terms and conditions.

EOI CONDITIONS :

8.1 The details of the medical equipments with specifications are mentioned in **Section VI. The firm must clearly mention their specification, special features, upgraded version (if any), detail technical catalogue of the offered model in their EOI.**

8.2 EOIs should be typewritten or computerized and every correction in the EOI should invariably be attested with signature by the EOIer with date before submission, failing which the EOI will be ineligible for further consideration.

8.3 Rates inclusive of excise duty / customs duty, packing, forwarding, insurance, transportation charges with **7 years onsite comprehensive warranty** and exclusive of Sales GST/TAX & Entry TAX should ***be quoted for the medical equipments (Item wise) on door delivery basis. The turnkey job (cost of accessories if any required for Installation/Commissioning including installation/commissioning charges), 7 year cost &*** Sales GST/TAX & ***Entry*** Tax ***should be mentioned in separate columns***. The rates quoted should be in **Indian Rupees only**. Rates quoted in any other currency will not be accepted.

8.4 The purchaser shall be responsible only after delivery and due verification, installation and commissioning of the equipment.

8.5 The rate per unit shall not vary with the quantum of order placed for destination point.



8.6 If there is difference between figures & words, words will be taken into consideration.

8.7 In the event of the date being declared as a holiday by Govt. of Chhattisgarh, the due date of sale, submission of bids and opening of bids will be the following working day at the scheduled place & time.

8.8 The price quoted by the EOIers shall not in any case, exceed the controlled price, if any, fixed by the Central / State Government and the Maximum Retail Price (MRP). The purchaser, at his discretion, will in such case, exercise the right of revising the price at any stage so as to confirm to the controlled price or MRP as the case may be.

8.9 The rate quoted and accepted will be binding on the EOIer for a period of **one year** from the **date of approval of the rate contract** and on no account, any increase in the price will be entertained till the completion of this EOI period.

8.10 No EOIer shall be allowed at any time on any ground whatsoever to claim revision of or modification in the rate quoted by him. Clerical error / typographical error, etc. committed by the EOIers in the EOI forms shall not be considered after opening of EOIs. Conditions such as " **SUBJECT TO AVAILABILITY"** / **"SUPPLIES WILL BE MADE AS AND WHEN SUPPLIES ARE RECEIVED**" etc., will not be considered under any circumstance and the EOIs of those who have given such conditions shall be treated as incomplete and for that reason, shall be rejected.

8.11 If at any time during the period of rate contract, the price of EOIed item is reduced or brought down by any law or act of the Central or State Government or the EOIer, the EOIer shall be morally and statutorily bound to inform the purchaser immediately about such reduction in the contracted price. The purchaser is empowered to unilaterally effect such reduction in rate, in case the EOIer fails to notify or fails to agree for such reduction of rate.

8.12 Approved rate with terms, conditions & the quoted price of the EOI shall remain valid for a period of **12 months from the date of approval**.



8.13 If the relevant documents / certificates which are required to be furnished along with the EOI are written in language other than English, the EOIing firm shall furnish English version of such documents / certificates duly attested by a Gazetted Officer / Notary with his seal and signature.

8.14 If any information or documents furnished by the EOIer with the EOI papers are found to be misleading or incorrect at any stage the EOI of the relevant items in the approved list shall be cancelled and steps will be taken to blacklist the said firm for three (3) years.

8.15 Rate should be quoted in Indian Currency, both in words and figures against each item as the payments will be made in Indian currencies only (Annexure-IX). The EOIer shall not quote his own rate for any item other than the item specified in the list. (**Section V – Schedule of Requirement**).

8.16 Both Cover-A and Cover-B should have an **index and page number** of all the documents submitted inside that cover.

8.17 The GST/TAX will be charged as per the guidelines given by the Finance Dept., Govt. of Chhattisgarh from time to time. Either GST/TAX (as applicable) will be paid to the supplier. In case of Tax, the supplier has to deposit the original receipt to claim it, if finished goods are brought from outside the State. The Sales Tax & entry tax components should be shown **separately** in the Price Schedule.

8.18 The requirement of items may increase or decrease depending on the situation.

PACKAGING :

9.1 All the packaging should be New. The supplier shall provide such packaging of the goods as is required to prevent their damage or deterioration during transit to their final destination. The packaging shall be sufficient to withstand without any limitation including rough handling during transit, exposure to extreme temperature, salt and precipitation during transit and upon storage.

COMPREHENSIVE WARRANTY & CMC :



11.1 The comprehensive warranty will remain valid for 3 years from the date of installation & commissioning of the equipment. The original copy of warranty documents will be submitted to the purchaser at the time installation.

11.2 The warranty will cover **all the parts of the machine or item and any replacement or repair required** within the warranty period and will be provided by the supplier free of cost at the destination point (installation point). The supplier will take back the replaced parts / goods at the time of their replacement. No claim whatsoever shall be on the purchaser for the replaced parts / goods thereafter. No traveling allowances or transportation cost will be paid by the purchaser during the warranty period.

11.3 The Supplier shall warrant that the Goods supplied under this contract are new, unused, of the most recent or current models and they incorporate all recent improvements in design and materials. The Supplier shall further warrant that all Goods supplied under this contract shall have no defect arising from design, materials or workmanship or from any act or omission of the Supplier that may develop under normal use of the supplied Goods in the conditions prevailing in the place of final destination.

11.4 **CMC**: The EOIer shall also commit to provide offer for CMC (**Labour + all spare**) for the **next 7(seven) years after five (3) years of warranty.** No extra cost will be paid other than the CMC cost for functioning of the item during this period. The supplier will provide one **(1) preventive** maintenance in every **six months** in a year during the period of CMC.

11.5 **The selected firm should have a service centre in CHHATTISGARH.**

11.6 All the warranty certificates must be handed over to the consignee at the time of installation.

TRANINING & OPERATIONAL MANUAL:

12.1 The firm / supplier will provide hands on training to two doctors and two technicians in his own cost for operating / handling the medical equipment(s) at the time of installation of equipment.



12.2 The supplier / firm will provide the operation / maintenance manuals of all equipments to the purchaser at the time of installation.

UPTIME GUARANTEE:

13.1 **UP-TIME BALANCE :**

The Supplier (s) shall provide guarantee 95% uptime during comprehensive warranty period, i.e., for 3 years from the date of installation & commissioning. Any uptime less than the specified period above will be compensated by the Supplier(s) by extending the warranty period. The consignee shall maintain a logbook in the format provided by the Supplier(s) which will indicate usage of the equipment every day and for calculation of up-time.

DOWNTIME PENALTY CLAUSE:

14.1 During the Guarantee / warranty period, desired uptime of 95% of 365 days will be ensured (24 hour). If downtime exceeds 5%, penalty in the form of extended warranty, double the number of days for which the equipment goes out of service will be applied. The supplier must undertake to supply all spares for optimal upkeep of the equipment for **TEN YEARS** after installation. If accessories / other attachment of the system are procured from the third party, then the supplier must produce cost of the accessory / other attachment and the CMC from the third party separately along with the main offer and the third party will have to sign the CMC with the purchaser if required. In no case equipment should remain in non-working condition for more than 7 (seven) days from the date of complaint, beyond which a penalty will be applicable as per Rule.

14.2 The principals or their agents are required to submit a certificate that they have satisfactory service arrangements and fully trained staff available to support the uptime guarantee.

SPARE PARTS:

15.1 The spare price list of all spares and accessories (including minor) required for maintenance and repairs in future after guarantee / warrantee period should be attached / enclosed along with the sealed quotation.



15.2 The EOIers are required to furnish the list of spares along with their cost in the financial Bid separately which will not be taken for evaluation.

15.3 Local agents / distributors quoting on behalf of the manufacturer / importer must attach the authority letter in their favour.

LABELLING :

16.1 The equipment supplied must be properly labelled with Sl. No., Model Name, Make & year of Manufacture

ACCEPTANCE OF EOI AND SUPPLY CONDITIONS:

17.1 The Purchaser reserves the right to reject the EOIs or to accept the EOIs for the supply of the item EOIed without assigning any reason thereof.

17.2 The Purchaser will be at liberty to terminate the contract either wholly or in part without assigning any reasons thereof. The EOIers will not be entitled to any compensation whatsoever for such termination.

17.3 The **supply should be completed within 30 days** from the date of issue of purchase order unless otherwise specified. If no supply is received even after 30 days with liquidated damage from the date of issue of the purchase orders , such orders will stand cancelled automatically without further notice. Penalties shall also thereafter be applied to the EOIer as specified in clause no. 21.1 to 21.2. The approved firm shall also suffer forfeiture of the EMD and Performance Security Deposit.

17.4 The EOI inviting authority or his authorised representative (s) has the right to inspect the factory of those company who have quoted for the EOI, before accepting the rate quoted by them or before releasing any purchase order (s) or at any point of time during the validity period of EOI and has also the right to reject the EOI or terminate / cancel the orders issued or not to reorder based on the facts brought out during such inspections.



EVALUATION:

18.1 The price bid of the EOIs who qualify in the technical bid fulfilling the eligibility criteria and complying to the technical specification shall only be opened.

18.2 The EOI inviting authority may ask for demonstration of the equipment by the bidders at the premises of the EOI inviting authority as a part of the technical evaluation before opening of price bid in order to verify the compliance to technical specification.

18.3 *The rates of t*he item quoted by the EOIer who qualify technically will be evaluated after taking the following points into consideration: -

a) Rate of the medical equipments will be taken after inclusion of the excise duty / customs duty, transportation, insurance, packing & forwarding & comprehensive warranty for (3) years, (cost of accessories if any for installation/commissioning and installation charges) & CMC for next seven years but excluding GST/TAX

b) The cost of the medical equipments (excise duty / customs duty, transportation, insurance, packing & forwarding & comprehensive warranty for five(Five) years but excluding GST/TAX), next seven(7) years after warranty will be added for evaluation.

19.1 If the supplier fails to complete the supply within the extended period, i.e. 30days after being allowed by the purchaser, no further purchase order will be placed to the firm for the said item including forfeiture of the Performance security and the

concerned firm will be blacklisted for two (2) years from the date of issue of letter for the said item.

TERMS OF PAYMENT :

20.1 No advance payments towards cost of medical equipments or turnkey job will be made to the EOIer.

20.2 Payments as mentioned above will only be made after keeping the **performance security deposit** from the supplier as per clause no. 7.1, if they have not deposited the same before. Payment will only be made after ensuring signing of the Agreement, undertaking and handing over of warranty papers of equipment and turnkey jobs by the supplier to the purchaser.

20.3 No claims shall be made against the purchaser in respect of interest on earnest money deposit or performance security deposit or any delayed payment or any other deposit.

20.4 Payments in shape of Draft / Pay Order will preferably be despatched to the supplier by Registered post

20.5 The payment of CMC will be made on a **six monthly basis**, after completion of warranty period and signing of the CMC agreement.

PENALTIES :

21.1 If the successful EOIer fails to deposit the required performance security within the time specified or withdraws his EOI after acceptance of his EOI owing to any other reasons or unable to undertake the contract, his contract will be cancelled and the earnest money deposit / performance security deposit submitted shall stand forfeited by the purchaser.

21.2 Violating the EOI terms and conditions & non supply / supply which is not as per technical specification will disqualify the firm to participate in the EOI for a period of 3 (three) years from the date of issue of letter and his E.M.D & performance security deposit will be forfeited and no further purchase order will be placed to that firm for that item.



21.3 In the event of any dispute arising out of the EOI, such disputes would be subject to the jurisdiction of the Civil Court of the concerned District or High Court of Chhattisgarh.

INSPECTION/TESTING :

22.1 The selected supplier shall have to arrange for demonstration of the equipment at the supply point. The purchaser or its nominated representative(s) shall inspect and test the equipments at the supply point to check their conformity to the technical specifications and other details incorporated in the contract.



ANNEXURE –I

(Refer Clause No. 3.1)

CHECK LIST

(To be submitted in **Cover A Technical Bid**)

Note : The documents has to be arranged serially as per the order mentioned in the check list

Please put ✓ in the respective box

COVER – A (TECHNICAL BID)	DOCUMENTS : SUBMITTED OR NOT		
1. List of Item (s) – Annexure II	Page No	Yes	No
2. EOI document Fee	Page No	Yes	No
3. Earnest Money Deposit	Page No.	Yes	No
4. Details of Manufacturing Unit / contact person Liaisioning agent / servicing centre (Annexure III)	Page No.	Yes	No
5. Declaration form (Annexure -III) signed by the tender & affidavit before Notary Public / Executive Magistrate	Page No.	Yes	No
6. Manufacturer's Authorization Format (Annexure –IV)(for distributor/Importer)	Page No.	Yes	No
7. Proof of avg. Annual turnover of Rs. 2.5 Crore Crore or more for preceding 3 financial years (for Manufacturer/Importer) or Rs.2 Crore or more (for authorized distributors) (Annual turnover for the manufacturer/importer is also to be submitted in case of distributor)	Page No.	Yes	No.
8. Audited Account Statemen (P&L Account) /AnnualReport for the three financial years by highlighting the figure in it, which is mentioned in the Annexure -V	Page No	Yes	No
9. Performance Statement **(Item wise)** during the last three year (Annexure -VI)	Page No.	Yes	No
10. Copies of Purchase order **(Item wise)** in support of the performance statement	Page No.	Yes	No
11. Deviation/No deviation Statement **(Item wise)** & details of technical specification	Page No	Yes	No



12. Leaflets/Technical Brocheures of the Equipment offered **(Item wise)**	Page No.		Yes		No.	
13. Copy of Import license (In case of Importer)	Page No.		Yes		No	
14. Copy of Valid ISO Certificate	Page No.		Yes		No	
15. Attested Photocopy of Up-to-date CE / US FDA/BIS Certificate (Item wise) (As per technical specification)	Page No.		Yes		No	
16. Attested Photocopy of Up-to-date IEC Certificate (Item wise) (As per technical specification)	Page No.		Yes		No	
17. Photocopy of PAN	Page No.		Yes		No	
18. Photocopy of Tax Clarence certificate	Page No.		Yes		No	
19. Copy of original EOI/tender and schedules, duly signed by the tenderer	Page No.		Yes		No.	



Annexure II

(Refer Clause No. 3.2)

(To be submitted in ***Cover A -Technical Bid***)

LIST OF ITEM(S) QUOTED

Sl.	Name of Item(s)	Name of Manufacturer	Make	Model Name	Details of offered product at Page No. (s)
1					
2					
3					

Signature of the tenderer :

Date :

Official Seal:

Annexure IIA

(Refer Clause No. 8 of Section -II)

(To be submitted in ***Cover A -Technical Bid***)

DETAILS OF EMD(S) SUBMITTED

Sl.No	Name of Equipment	EMD Amount (Rs.)
	TOTAL (Rs.)	

Signature of the tenderer :

Date :

Official Seal:

ANNEXURE – III

DECLARATION FORM

(On Non Judicial Stamp Paper worth Rs. 100/-)

I/We……..having My / our office at……..……..……..do declare that I / We have carefully read all the terms & conditions of EOI of Ref. No.……., of CS CUM HS DISATRICT HOSPITAL KONDAGAON., for the supply of ……... and confirm our eligibility for this EOI and all items quoted as per the EOI condition and Governing laws of India.

We certify that the rates of items quoted are reasonable & not higher or the prices charged by us to Wholesalers or for institutional supplies in last six months.

I/We do hereby declare all information provided is true. I/We do hereby declare I/We have not been convicted by any court of Law nor I/We are de-recognized / black listed by any State Govt. / Union Territory - Govt. of India / Govt. organization / Govt. Health Institutions for supply of Not of Standard Quality (NSQ) items / part-supply / non-supply/non-service or any reason whatsoever. I/We agree that the EOI Inviting Authority can forfeit the Earnest Money Deposit and or Security Deposit and/or blacklist me/us for a period of 5 years if, any information furnished by us proved to be false at the time of inspection / verification and not complying with the EOI terms & conditions.

I / We …… do hereby declare that I / we will supply the items as per the terms, conditions & specifications of the EOI document

Signature of the bidder:

Date:

Name & Address of the Firm:



ANNEXURE – IV

(On 100Rs. Stamp Paper)

FORMAT FOR SUBMITTING LIST OF ITEM FOR WHICH BID IS QUOTED

Sl. No.	Item Code	Name Of the ITEM	Make, Model and Manufacturer	Address of Manufacturing Premises/Unit
1				
2				
3				
4				

Total EMD Paid:

UTR / RTGS Transaction No:

I/We……..having My / our office at……..…….…….. declares that I / We will supply the items as per above.

Sign & Seal of the bidder.

ANNEXURE – V

TOTAL TURNOVER CERTIFICATE

To

CIVIL SURGEON CU HOSPITAL SUPRITENDENT

DISTRICT HOSPITAL KONDAGAON Chhattisgarh

We hereby certify that **M/s**_____ (the name of participant in the EOI) who is participating the EOI for supply of Goods called by CS CUM HS DISTRICT HOSPITAL KONDAGAON Chhattisgarh, having their office at_____(Address of office) has a turnover on their audited balance sheet as follow :-

Turnover in the financial year of 2013-2014.	RS.
Turnover in the financial year of 2014-2016.	RS.
Turnover in the financial year of 2015-2016.	RS.

The above information is correct and true and verified from audited Balance Sheet.

CHARTERED ACCOUNTANT

(with membership no.)

Name:

Contact No.:

Contact Address:

NOTE: 1 The turnover of other than participant will not be accepted.



ANNEXURE – VI

List of Govt. Organization to which bidder is a supplier

(Supplies within last 3 years)

Name of the Bidder/ Supplier:

Sl. No	Order placed by (Full Address of Purchaser) (1)	Order No. and Dated (2)	Description and Quantity of ordered equipment	Value of order	Date of Completion of delivery	Remarks indicating reasons for late delivery, if any
1						
2						
3						
4						
5						

Signature and Seal of the Bidder.....

.....

Note: Format may be used in landscape, certificates from the purchasers in the format enclosed.

ANNEXURE – VII

DETAILS OF THE BIDDER& EOI CONTACT PERSON

	Corporate Office (The address in which the purchase orders and payment details will be communicated)	Contact For this EOI/Tender
Name & Full Address		
Telephone Nos., Landline		
Mobile		
Fax		
E – Mail		
Date of Inception		
Manufacturing / Import License Nos. & Date		
Name of the issuing Authority		
License valid up to		

Signature of the tenderer:

With seal



ANNNEXURE – VIII

Bank Mandate format :-

(in lieu of the bank certificate to be obtained , please **attach the original cancelled cheque**issued by your

01	**Name of the Bank .** **Branch Name& address.** **Branch Code No.** **Branch Manager Mobile No.** **Branch Telephone no.** **Branch E-mail ID**	
,	**9 digit MICR code number of the bank and branch appearing on the MICR cheque issued by the bank.**	
03	**IFSC code of the Branch**	
04	**Type of Account (Current / Savings).**	
05	**Account Number (as appear in cheque book)**	

bank for verification of the above particulars).

I /We hereby declare that the particulars given above are correct and complete. If the transaction is delayed or not effected at all for reasons of incomplete or incorrect information, I would not hold CS CUM HS DISTRICT HOSPITAL KONDAGAON responsible. I have read the conditions of the EOI/agreement entered and agree to discharge the responsibility expected of me / from the company as a tenderer /successful tenderer.

Date: Company Seal Signature

Place: (Name of the person signing & designation)

CERTIFIED THAT THE PARTICULARS FURNISHED ABOVE BY THE COMPANY ARE CORRECT AS PER OUR RECORDS.

Bank Seal with address. Signature of the authorized official of the bank.

---.

ANNEXURE – IX

PRODUCT SPECIFICATION

(To be uploaded in Cover B: Technical Bid)

Item Code:

Name/ Description of Equipment:

A	B	C	D
Desired Specification of TIA	Manufacturer's Specification	Deviation	Remark

Note: This form is product/Equipment specific

- Desired Specification should be taken from Annexure – II (Specification of equipments)
- Manufacture/Bidder should quote their specification in Column 'B'.
- Column 'C' and 'D' should kept blank for evaluation purpose.

Signature and Seal of Bidder

Format may be used in Landscape.

Annexure VIIIA

(*Refer Clause No. 3.10*)

STATEMENT REGARDING DEVIATIONS FROM TECHNICAL SPECIFICATIONS (IF ANY)

Following are the Technical deviations and variations from the purchaser's Technical Specifications.

Sl. No.	Item Name	Clause of Technical Specification Statement	Deviations/ Variations if any
1			
2			
..			
..			
..			

In case there is no deviation from technical specification, Pl. Mention ***No Deviation.***

Signature of the Bidder

Name :
Date :

Place :

Seal

(To be submitted in ***Cover A -Technical Bid***)

Annexure VIII B

(***Refer Clause No. 3.10***)

DETAILS OF TECHNICAL SPECIFICATION OF THE PRODUCT (S) OFFERED BY THE BIDDER

Sl. No.	**Item Name**	**Make**	**Model**	**Detail Specification of the product(s) offered* (Pl. Describe the detail specification of the product offered) – Para wise compliance to the technical specification asked for.**	****Page no. of the Catalogue / Leaflet where Para wise compliance information as per technical specification is available**
1					
2					
..					
..					

* Leaflets/Technical Brocheures of the product offered must be attached in support of the information provided above.

** It is mandatory to mention the page no(s) in the format as mentioned above.

Signature of the Bidder
Name :

Date :

Place :

Seal

To be submitted in *Cover B – Price Bid*

(Refer Clause No. 4.1 & 8.15)

FORMAT I - PRICE SCHEDULE

Name of the Item (s) (Items mentioned in the schedule of requirement)	**Make & Model**	***Price** of the **items** with **all accessories** which includes excise duty / customs duty, packing, insurance, forwarding / transportation (door delivery) with **3 (three) years onsite warranty, calibration charges if any** & **excludes GST/TAX** Cost in **Rupees** (both in words & figures)	CMC (excluding **Service Tax**) for five years **after** expiry of five years warranty period (please mention on yearly basis)	**Cost of **if any** (all materials : for electrical, civil , air conditioning work etc.) for installation & commissioning and all charges towards installation/commissioning including **all taxes for** in Rs. (Door delivery & installation)	Total cost	GST/TAX	
(1)	**(2)**	**(3)**	**(4)**	**(5)**	6	**7**	**8**
			1st year after warranty: 2nd year after warranty: 3rd year after warranty: Total :				

* **Breakup of the price of individual items of the items mentioned at col. (3) above should be mentioned separately at Annexure**

** **The cost of turnkey shall include any specific accessories/equipment required for installation/commissioning and . In case of turnkey, the details of accessories/equipment are to be mentioned.**

*** **GST/TAX which will be chargeable on the price (3) shall be mentioned separately in column 7 above.**

Date :

Place :

1. **Rates should be quoted both in figures & words for each item and if there is any discrepancy, the quoted rates in words will be taken for evaluation.**
2. **The EOIer has to mention the make / brand, specification, warranty of all the items in turn key.**

GST/TAX paid to Government of CHHATTISGARH



ANNEXURES

(Agreement, Warranty and CMC Undertaking)

ANNEXURE – X

(Refer clause no. 7.1.2)

AGREEMENT

THIS AGREEMENT IS MADE AT _____ THIS THE DAY OF _____ **2017**

BETWEEN

Name of the Supplier
with full address

Here in after called the "Supplier(s) _____" as 1st Party

AND

Civil Surgeon cum Hospital Superitendent District Hospital Kondagaon Chhattisgarh

herinafter called the "PURCHASER" _____ as 2nd Party.

Relying on the documents and representation of facts connected to the issue of aforesaid parties to undertake the responsibilities of sell and purchase of following equipment(s) etc. with the terms & conditions hereinafter laid down.

And whereas the 2nd party "Purchaser(s)" is willing to purchase

Name of the Item:

Specifications: As per specifications laid down in the EOI terms & conditions

The Supplier(s) has agreed to sell the equipment(s) completed in all respects according to the EOI requirements and their / his offer dtd._____ and the Supplier(s) has also agreed to install to make them operative at the destination mentioned in the EOI document with the following descriptions and their cost mentioned against each.

Description of goods	Qty	Price	Total

The price / cost of the item also include the followings in addition to above.

1. Insurance
2. Freight
3. Transportation
4. Customs duty / Excise duty
5. Charges for documents, instructions manual, tools
6. F.O.R. at the destinations mentioned in the consignee list
7. Training to doctors & technicians.

1. Maintenance of the system includes all accessories supplied and their spare parts required during comprehensive warranty period of five year at free of cost from the date of successful installation and satisfactory functioning of the system at the site.
2. Installation and commissioning of the system by the Supplier's engineer at site.
3. Any other charges including loading & unloading, packing & forwarding etc. will be paid by the Supplier(s) till the completion of the installation and turnkey job if any.

CMC cost for next 7(seven) years after the warranty period shall be paid after completion of the warranty period (on a six monthly basis).



TERMS AND CONDITIONS:-

PRICE :

Only the price quoted by the Supplier(s) in his / their financial proposal will be the price for payment and no other price escalation will be allowed at any circumstances.

SUPPLY

The supply should be completed within 30 days from the date of issue of purchase order unless otherwise specified. If no supply is received even after 30 days with liquidated damage from the date of issue of the purchase orders , such orders will stand cancelled automatically without further notice. Penalties shall also thereafter be applied to the EOIer as specified under Penalty. The approved firm shall also suffer forfeiture of the EMD and Performance Security Deposit.

LIQUIDATED DAMAGE :

If the supplier fails to complete the supply within the extended period, i.e. 30 days after being allowed by the purchaser, no further purchase order will be placed to the firm for the said item including forfeiture of the Performance security and the concerned firm will be blacklisted for two (2) years from the date of issue of letter for the said item.

TERMS FOR PAYMENT :-

A. The payment(s) shall be made by purchaser in Indian currencies. No advance payments towards cost of Instruments and Equipments etc. will be made to the EOIer.

B. Before release of payment the supplier has to submit the signed agreement, warranty documents of equipment and turnkey job to the consignee. The undertaking as per Annexure – XI & XII will also be submitted to the consignee with photocopies to the purchaser.

C. The payment of CMC will be made on six monthly basis after expiry of the warranty period and signing of the CMC agreement.



DOWNTIME PENALTY CLAUSE:

During the Guarantee / warranty period, desired uptime will be 95% of 365 days (24 hour) if downtime exceeds 5%, penalty in the form of extended warranty, double the number of days for which the equipment goes out of service will be applied. The vendor must undertake to supply all spares for optimal upkeep of the equipment for TWO **YEARS** from the date of installation at the site. If accessories / other attachment of the system are procured from the third party, then the vendor must produce cost of accessory / other attachment and the CMC from the third party separately along with the main offer and the third party will have to sign the CMC with the consignee if required.

In no case equipment should remain in non-working condition for more than 7 working days.

The manufacturers or their agents are required to submit a certificate that they have satisfactory service arrangements and fully trained staff available to support the uptime guarantee.

COMPREHENSSIVE MAINTENANCE CONTRACT:

The supplier will provide CMC for 7 (seven) years after the completion of 3 years comprehensive warranty period.

INSTALLATION AND DEMONSTRATION :

The installation and demonstration of the equipment shall be done by the Supplier(s) at free of cost at the installation site of the respective institutions.

TRAINING :

Supplier(s) shall impart adequate training to 2 doctors and 2 technicians at the site / his / their factory / workshop inside / outside India as the case may be at the Supplier(s) cost.

INCIDENTAL SERVICES :

The Supplier(s) shall abide by the terms and conditions under incidental services & the installation of Instrument / Equipment at the destination point (Door Delivery) of consignee and demonstrate the machine in working condition to the receiving authority.

Furnishing of tools required for assembly and / or maintenance of the supplied Instruments / Equipments.

Furnishing of detailed operations and maintenance manual literatures for each appropriate unit of supplied Goods.

Performance or supervision or maintenance and / or repair of the supplied Goods, for a period of five (5) years i.e. the warranty period, provided that this service shall not relieve the Supplier of any warranty obligations under this contract.

The successful supplier shall replace any part or whole system as may be necessary in the event of damage during transit or found damaged on arrival or during installation of the system or if found not in conformity to the specifications at his / their own cost.

The EOIer should furnish an undertaking to the effect that he / they should take responsibility after sales service of the equipments / instruments to be supplied by him / them and to provide spare parts for up keeping the Equipments / Instruments for a minimum period of 10 years from the date of installation.

The price of the instruments / equipments is inclusive of warranty for a period of 5 (five) years commencing from the date of installation. The EOIers shall submit undertaking for C.M.C (Comprehensive Maintenance Cost) for a period of 7 (seven) years from 8th year onwards duly signed by authorised signatories for the execution at appropriate time).

SPARE PARTS :

The supplier will provide all the spare parts, repairing & maintenance by its trained personnel after the warranty period (5 years) during the CMC period.

COMPREHENSIVE WARRANTY :

This warranty shall remain valid for three (3) years from the date of installation & commissioning of the machine / item & must be submitted at the time of installation to the consignee with a photocopy to the purchaser.

The warranty will cover all the parts of the machine or item and any replacement or repair required within the warranty period will be provided by the supplier free of cost at the destination point (Installation point). The supplier will take back the replaced parts / goods at the time of their replacement. No claim whatsoever shall be on the purchaser for the replaced parts / goods thereafter. No traveling allowances or transportation cost will be paid by the purchaser during warranty period.

The Supplier warrants that the Goods supplied under this contract are new, unused, of the most recent or current models and they incorporate all recent improvements in design and materials (even if the advanced facilities are not mentioned in our product specification). The Supplier further warrants that all Goods supplied under this contract shall have no defect arising from design, materials or workmanship (except when the design and / or material is required by the Purchaser's Specifications) or from any act or omission of the Supplier, that may develop under normal use of the supplied Goods in the conditions prevailing in the place of final destination.

The Purchaser / consignee shall promptly notify the Supplier in writing / Fax / Telephone of any claims arising under this warranty.

Upon receipt of such notice, the Supplier shall with all responsible speed will repair or replace the defective goods or parts thereof without cost to the purchaser to maintain its UP TIME offered in the beginning of purchase otherwise penal provisions shall apply if the supplier fails to keep up its UP TIME.

If the Supplier, having been notified, fails to remedy the defect(s) within 10 days, the Purchaser may proceed to take such remedial action as may be necessary, like forfeiture of EMD or recovery from security deposit the amount of loss incurred by the purchaser.



GOVERNING LANGUAGE :

The contract shall be written in English language. English language version of the contract shall govern its interpretation. All correspondences and other documents pertaining to the contract which are exchanged by the parties shall be written in English.

DELIVERY OF DOCUMENT :

Four (4) copies of the Supplier invoice / bills showing purchase order number, good's description, quantity, unit price, total amount with stock entry certificate by the consignee.

Photocopy of the Insurance Certificate if any (The Original Certificate is to be given to the Consignee).

Attested Photocopy of Manufacturer's / Supplier's warranty certificate. (The original warranty certificate is to be submitted to the consignee at installation point).

INSURANCE :

For delivery of goods at site, the insurance shall be obtained by the Supplier(s) in an amount equal to 110% of the value of goods from "Warehouse" (final destination) on "All Risks" basis including natural calamities.

PACKAGING :

The supplier shall provide such packaging of the goods as is required to prevent their damage or deterioration during transit to their final destination. The packaging shall be sufficient to withstand without limitation rough handling during transit and exposure to extreme temperature, salt and precipitation during transit and upon storage. All primary packaging containers which come in contact with the item should strictly protect the quality and integrity of the Instruments & Equipments. Packing case size and weights should be taken into consideration, in case of remoteness of final destination and the absence of heavy handling facilities at all points in transit.

The packaging marking shall show the description of quantity of contents, the name of the consignee and address, the gross weight of the packages, the name of the supplier with a distinctive number of mark sufficient for purposes of identification. Each package shall contain:

i. a packaging note quoting the name of the purchaser
ii. the number and date of order
iii. nomenclature of the goods
iv. schedule of parts for each complete equipment giving part number with reference to assembly.
v. Name & address of the consignee
vi. Name & address of the supplier.

TERMS OF CONTRACT :

The **Civil Surgeon cum Hospital Superitendent District Hospital Kondagaon Chhattisgarh** will be at liberty to terminate the contract either wholly or in part without assigning any reason. The EOIers will not entitled to any compensation whatsoever in such terminations.



PENALTIES :

If the successful EOIer fails to execute the agreement and / or deposit the required security within the time specified or withdraws his EOI after acceptance of his EOI owing to any other reasons, he is unable to undertake the contract, his contract will be cancelled and the Earnest Money Deposit deposited by him along with his EOI shall stand forfeited and he will also be liable for all damages sustained by the **Civil Surgeon cum Hospital Superitendent District Hospital Kondagaon Chhattisgarh**
by reasons of such breach, such as failure to supply / delayed supply including the liability to pay any difference between the prices accepted by him and those ultimately paid for the procurement of the articles concerned. Such damages shall be assessed by the **Civil Surgeon cum Hospital Superitendent District Hospital Kondagaon Chhattisgarh** whose decision is final & binding in the matter.
If any articles or things supplied by the EOIer have been partially or wholly used or consumed after supply and are subsequently found to be in bad order, unsound, inferior in quality or description or are otherwise faulty or unfit for consumption / use & rusted then the contract price or prices of such articles on full will be recovered from the EOIer, if payment had already been made to him or the EOIer will not be entitled to any payment for that item & no further order will be given to him. For infringement of the stipulations of the contract or for other justifiable reasons, the contract may be terminated by the **Civil Surgeon cum Hospital Superitendent District Hospital Kondagaon Chhattisgarh** and the EOIer shall be liable for all losses sustained by the **Civil Surgeon cum Hospital Superitendent District Hospital Kondagaon Chhattisgarh**
in consequence of the termination which may be recovered from the Security Deposit made by the EOIer or other money due or become due to him.

Supply of sub-standard items or non - performance of EOI terms & conditions will disqualify a firm to participate in the EOI for the next five years.

ARBITRATIONS :

In the event of any dispute out of the contract, such dispute should be subject to the Jurisdiction of the High Court, chhattisgarh

CHANGE OF TERMS AND CONDITIONS :

Any amendment to the terms & conditions and clauses of the agreement if required must be done in writing duly signed by the two parties.

IN WITNESS WHERE OF the parties herein to have set and subscribed their respective hands the day and year first herein above written.

Executed by Purchaser (s) / Consignee

Executed by Supplier(s)

In presence of (Witness)

In presence of (Witness)



ANNEXURE – XII

(Refer Clause No. 11.1 to 11.6 & 13.1)

UNDERTAKING

(To be submitted on Rs.50/- stamp paper)

EOI ref. No. _____

Name of the equipment:

Date of Installation:

Name of the Consignee:

Name of the purchaser:

Sir,

I / we _____

hereby declare that

1. I / we am / are the manufacturers / authorized agents / distributors of _____ _____.
2. I / we do accept / agree for the all clauses including the warranty 3 years followed by 7 years CMC) and payment terms and conditions of this EOI.
3. I / we do hereby confirm that the prices / rates quoted are fixed and are at par with the prices quoted by me / us to any other Govt. of India / Govt. of Chhattisgarh Hospitals / Medical Institutions. I / we also offer to supply the stores at the prices and rates not exceeding those mentioned in the price bid.
4. I / we agree to abide by my / our offer for a period of 365 days from the date of approval of the EOI.
5. I / we have necessary infrastructure for the maintenance of the equipment and will provide all the accessories / spares as and when required.
6. I / we also declare that in case of change of Indian Agent or for any other change, merger, dissolution solvency etc. in the organization of our foreign principles, we would take care of the Guarantee / warranty / maintenance of the machinery / equipment and have provided written confirmation for the same.
7. I / we shall provide assistance to the consignee in clearance and delivery of store at consignee's stores / premises.

8. The demurrage / storage charges, if any, payable to the customs department, due to non-receipt of required documents in time by the hospital / delay due to incorrect entries, mistakes to the documents etc. shall be borne by me / us.

9. I / we have carefully read and understood all the terms and conditions of the EOI and shall abide by them.

10. I / we undertake to get the equipment's repaired within 48 hours of receiving of the complaint from the indenting hospital / consignee failing which a penalty @ 1% of the cost may be recovered from the performance security before releasing the same to us after 5 years warranty period.

Signature of the witness
Name & address

Signature of the Tederer
Name & address

Dated

Seal of the firm.

N.B: 1. To be attested by Notary Public

2. Only to be submitted by the approved supplier / tenderer to the consignee and a copy to the purchaser before release of payment.

OFFICE OF CIVIL SURGEON CUM CHIEF HOSPITAL SUPERINTEDENT
DISTRICT HOSPITAL KONDAGAON (C.G.)

Expression of Interest

Expression of Interest is Invited is Seated Cover From Repeated and Interested Manufacture firm, Authorized Dealer for Supply Institution Medical Equipment and Furniture and Maintain Sealed Cover of FOI will be Received on or before 06 September 2017, Application fee Rs. 1000/- Demand Draft Jeevan Deep Samiti District Hospital Kondagaon Received in office of Civil Surgeon Cum Chief Hospital Superintendent District Hospital Kondagaon. For Detail Interest Dealer /Supplier/Authorized may Contact with Purchase Section of Department or Same can download from **District Website Kondagaon.gov.in** or State Website **cgstate.gov.in**

Civil Surgeon Cum Chief Hospital Superintendent
District Hospital Kondagaon (C.G.)

Ac and Refrigerator Equipment

S.No.	Item	specification
1	space ice pack Box	insulated ice box 50 liter Brand-Aristo,cello
2	Room Heater/Cooler for immunization clinic with electric fitting	
3	Wast disposal twin bucket,Hypoclorite solution	
4	thermometer Alchohal(Steam)	Contactless temperature measurement • Best usable for cold Chain equipment (WIC, WIF, ILR, DF, Cold Box, Vaccine Carrier) • Easy to hang (Horizontal) • Automatic shut off after a minute • Visible in red color • Safe for daily use • Manual
5	Almirah for Vaccine Logistic	Size: 960 × 400 × 1750 mm. Epoxy coated mild steel framework. Stainless steel bottom.
6	Almirah for Vaccine Logistic	Size: 960 × 400 × 1750 mm. Epoxy coated mild steel framework. Stainless steel bottom.
7	Immunization table	
8	Chair for New staff proposed	Colour: Black Seat Size: 21" W X 19" D Back Size: 19" W X 23" H Overall Dimensions: 24" W X 19" D X 34" - 36" H Material: Fabric Pneumatic Seat Height Adjustment. Medium Height Back Rest. Lumbar Adjustment on Back Small Cushion Pillow Support on Back Nylon Netted Mesh Back Adjustable Up-Down Arm Rest. Virgin Quality PP Stand/Base Gas lift Height Adjustment. 90 Degree Tilt Lock. Symcro Push Back Design Machanism. Imported Castor Wheels with Good Grip. 360 Degree Revolving.

S.No.	Item	specification
9	Stool for Immunization room	Adjustable Size: 320*460*620 Material: Metal Metal Type: Iron Folded: Yes
10	Bench for waiting Area	3 seater, Steel Size: 1900x630x900mm colour-blue Resistance against water Flat edges Highly very spacious (Airport three seater chair)
11	Dustbin with lid	animated Design , Made of High Density Polyethylene (HDPE) material Injection moulded biomedical wast standard High resistance to Heat and radiation UV stabilized Environment friendly and easy to clean Available in various colours Specifications: Nom. volume (L) 80 Dead weight approx. (kg) 9 Useful load (kg) 45 A - Overall height (mm) 780 B - Overall width (mm) 480 C - Overall depth (mm) 560 D - Upper edge comb (mm) 735 E - Wheel diameter (mm) 200 colour-green
12	water container	Material: Stainless Steel,Polypropylene And Polyurethane Insulation Water Storage Size: 10 Ltr Special Feature: Insulated With Top Quality Polyurethane ,Preserve The Hot Water Warm For Hours And Water Cold For Hours Warranty: 6 Months

S.No.	Item	specification
13	Hub cutter	Capacity 3 Liter Body made of Stainless Steel / ABS Upper Lid made of Virgin Plastic Syringe Hub & Needle gets cut off by a sharp blade by pulling the handle. Separate holes for inserting Needles & Syringe hub. no electricity required Economical & Reusable
14	coolers	Cools up to 2500*sq.ft. 30 inch Aluminium Fan 90 ltr.Tank (Seperated Tank) Aluminium & Stainless Steel Componentry Constructed by Injection Moulding Machine Portable,Ductable,Wall Mounted, Industrial,Evaporative Air Cooler Cooling Media - Honey Comb Pad with Cellulose Material RPM/Wattage/Hp - 950 RPM/490* W/0.67* Hp Power Consumption - 1 Unit in 2 Hours Gross Weight- 49kgs Air Flow(M3/H)/CFM - 18000* CMH/10500* CFM Cooling Pad Size -1100*100*1150mm Dimension - 1) 1135*700*1790mm/LOUVER 2) 1135*600*1790mm/GRILL
		Total

A. Child Health Equipment

S.No	Item	Specification
1	Resuscitator, Hand- Operated, neonate and child, 50 ml	
2	Laryngoscope Set- Pediatric Size	Complete with Standard Size Battery Handle made of Stainless Steel + Three Miller Stainless Steel Blades all Satin/Dull Finish and packed in Zip Case .
3	Electronic Baby Weighting Scale	Standard Capacities 20 kg Platter Size 295 x 500 mm abs Power Supply Main / Battery Display Digit 6 Digit Number of Keys 3 keys Battery Charger yes Battery yes Remote Control No Number of characters on Display 6 Digit Color Red/Green Storage Capacity Yes Outer Body abs Message Languages Numric Display 6 Digit (front) Function Power On, Power Off, Tare Power Mains/Battery (6V/4Ah) Printer No
4	Suction Pump, Foot operated	
5	Thermometer, Clinical digital	Clinical Thermometer Oval flat type Lens front, dual scale Cylindrical bulb Range: 35-45°C & 95-108°F Accuracy: ± 0.1C & ± 0.2°F In plastic case
6	Light For Examintaion	Technical Data of led Cold-Light Source : Lamp: led Color Temperature: 6000 KDevice type : I/B Lamp Power :Equivalent to 300Watt xenon Input Power : 400 W Case Temperature : <55 °C Working Noise : <55DB Operating Life : 50,000.00 hours Outside electrical source: 220V±10%,50Hz Working Environmental Humidity : 5-32 °C Relative Humidity : <70% Dims. : standard Net.Weight. : standard
7	Syringe Hub Cutter	

S.No	Item	Specification
8	Phototherapy unit, single head, high intensity	Irradiance: Minimum 40 µW cm-² nm ¹(+10± %) at 30 cm from the With adjustable Intensity (High, Medium & Low) Optical Sensor / Radiometer Measurement band 0 µW cm-² nm to 200 µW cm-² nm Resolusion- 0.1 µW cm-² nm Reading accuracy - ±2 µW cm-² nm Noise - <52 dBAWavelength>450-460 nm (peak 451 nm). free of UV and IR radiation Electrical Specification- Output Voltage :- 12 VDC Output power:- 50 W Input voltage:- 120 -240 V, 1.5 A Operating frequency:- 50/60 Hz Environments Requirement:- temperature range Operating 18 to 28 °C (64.4 to 82.4 °F) Storage -20 to 60 °C (-4 to 140 °F) Relative humidity Operating -10 % to 95 % (non condensing) Storage -10 % to 95 % (non condensing) atmospheric Pressure Operating -700 hPa to 1060 hPa Storage -500 hPa to 1060 hPa Data transmission RS-232 serial port -Standard UART protocolBits per second -9600 bps Data bits -8 Parity -None Stop bits- 1 Flow control- None Lamp type: Light emitting diode (LED) for 30,000 hours providing the same intensity level. Height : Adjustable from 1270 ± 20 mm to 1730 ± 20mm up to 90°on either side. Time totalizer: Digital (compact) 1. Time totalizer for the LED running time (non-resettable) 2. Therapy timer for the treatment time (resettable) Examination lamp: The LED PT is provided with 2 useful examination lamp that provides normal light when the baby needs to be attended to by doctors or nurses. Alarms:. 1. Cooling fan failure 2. Temperature unit in light unit exceeds 50°C 3. Internal supply voltage is high 4. Internal supply voltage is low 5. Current consumption of LED's is not normal (Messages are displayed on the time totalizer display. Audible alarm is also produced.)
		µ - light unit.Cooling fan: Centrifugal type to dissipate the heat created by LED's Power Requirements:-120 -240 V ~50/60 Hz , 1.5 A Epoxy/powder coated body for scratch and rust prevention and PU (polyurethane) coating for plastic parts. 3 swivel casters, including 2 with brakes

S.No	Item	Specification
9	Resuscitator, Hand- Operated, neonate and child, 200 ml	Resuscitator chilled 1 pcs Suction Foot Type Small 1 pcs L-Scope 4 Blade S.S. 1Set L-Scope 2 Blade S.S. 1Set Oxygen Hood Round Jointless 1Set E.T.Tube for Ped 3 Pcs Airways plastic (Set of 3 pcs) 1Set Styllete (Set of 3 pcs) 1 SetFace mask Round shape 1 Set Face mask rubber 1 Set Mucus Extractor 2 Pcs M-Forceps (Set of 3 pcs) 1 Set M-Brush(Set of 3 pcs) 1 Set Mouth Bite (Set of 3 pcs) 1 set
10	Suction Pump, Portable	
11	Surgical Instument, Suture / Set	
12	Oxygen Hood, S and M size, set of 3 each including connecting Tubes	
13	Oxygen Concentrator	1.To deliver 8 Litres /minDual Port (each can deliver 4 litres / min) Advantage: can be connected to 2 patients at a time 3. 24 hour usable 4. Flow rate – 1-10lpm 5. Outlet pressure – 18-20 psig Alarms – 1. Power failure, 2. High & Low pressure, 3. Low purity test, 4. temperature Advantage: can be connected to 2 patients at a time 6. Advantage: can be connected to 2 patients at a time
14	Pulse Oxymeter, bedside, neonatal	1. Method: Dual wavelength LED 2. Oxygen saturation Range 0-99% Resolution 1% Accuracy + 2 at 70-99% range, + 3 at 40-70% range Averaging time selectable (3-21 secs) Search time selectable (10-40 sec) Sensors Re-usable finger probe, Re-usable flexible multi-site probe 3. Pulse rate Waveform Plethysmographic or bar form Range 20-250 bpm Resolution 1 bpm Accuracy + 2 bpm 4. Display Saturation Pulse Rate Waveform Alarms4. Alarms High SpO2 Range 70-99% Low SpO2 Range 50-99% High pulse 20-250 bpm Low pulse 20-200 bpm System alarmsProbe failure Poor signal Power failure 5. Trends Memory 12 hours Data interval 20 sec Display 2-12 hours 6. Power 220/240 V AC, 50/60 Hz Rechargeable internal battery

S.No	Item	Specification
15	neonatal Stethoscope 1 for each Bed	Machined Stainless Steel · Chestpiece Technology - Double sided · Chestpiece Weight - 40 Gram · Diaphragm Diameter - 1.31 in / 3.3 cm · Diaphragm Material - Epoxy/Fiberglass · Diaphragm Type - Floating Diaphragm · Eartip Type - Soft Sealing · Extra Eartips - Yes · Headset Material - Wide Diameter Aerospace Alloy / Anodized Aluminum · Length - 28 in / 71 cm · Model - Pediatric · Net Weight (approx.) - 105 Gram
16	Sphygmanometer, neonate electric	Wall mounted B.P. Apparatus-Accuracy: +/-3mmHg · Sub-division: 2mmHg · Dial-type, mercury-free manometer · Measure scope: 0-300mmHg
17	Light Examination for mobile	
18	Measuring Tape, vinyl coated	
19	Kidney basin , stainless steel 825 ml	
20	Dressing tray, stainless steel	
21	Infusion stand, double hook	
22	Indicator, TST control spot	
23	Irradiance meter for phototherapy	
24	.Infantometer	Readability: to nearest mm. For measurement of length, for growth assessment in infants Technical Data Measuring range in cm:33 - 100 cm Measuring range in inch:13 -39 " Graduation [Measuring Rod]:1 mm / 1/16 inch Measure (W x H x D):1103 x 169 x 402 mm , 43,4 x 6,7 x 15,8 inch approx Net weight:3,8 kg , 8,3 lbs Functions: lying measurement
25	Bench top steam Autoclave electical	
26	Sterilising drum, 165 mm diameter	
27	Electic Steriliser	
28	Gowns for Staff and mothers	staff Gown -Dark blue colour,mother gown -purple colour
29	Washable Slippers	paragon small/medium/large
30	Haematocrit Centrifuge	
31	Binocular Microscope with IIIuminator	

S.No	Item	Specification
32	Bilirubinometer	Battery Number of measurements (when fully charged) Internal NiMH 250 AC adapter Input 100 V ~ to 240 V ~, 50/60 Hz, 11 VA to 18 VA Output 9 VDC, 500 mA Light source Pulse xenon arc lamp Light source life 150,000 measurements Sensors Silicon photodiodes PHYSICAL SPECIFICATION Width 56 mm Depth 45 mm Height 168 mm Weight 203 g ± 10 % PERFORMACE SPECIFICATIONS Measurement range 0.0 mg/dL to 20.0 mg/dL (0 µmol/L to 340 µmol/L) Clinical Data Standard Error of Estimate (SEE) ± 1.5 mg/dL or ± 25.5 µmol/L (> 35 weeks gestation) ± 1.6 mg/dL or ± 27.4 µmol/L(24 – 34 weeks gestation) DATA TRANSMISSION USB port HL-7 or CSV AMBIENT CONDITIONS DURING OPERATION Temperature 10 °C to 40 °C (50 °F to 104 °F) Air pressure 700 hPa to 1060 hPa Altitude range -400 m to 3.000 m Relative humidity 30 % to 95 % (without condensation) DURING STORAGE AND TRANSPORTTemperature -10 °C to 50 °C (14 °F to 122 °F) Air pressure 700 hPa to 1060 hPa Altitude range -400 m to 3.000 m Relative humidity 30 % to 95 % (without condensation)
33	Glucometer with Dextrostix	
34	Refrigerator 110L	5 star LG with Stablizer two chember door
35	Voltage Servo Stabiliser	
36	Room heater	
37	Spot Lamps	
38	Wall Clock with Second Hand	

Furniture & Hospital Accessories for		
Sl. No.	**Name of the Equipment**	**specification**
1	Doctor's chair for OP Ward, Lab etc.	Chair Ambulatory is made of Aluminium Round Tube The seat and back are foam padded It comes with padded arm rests Aluminum frame structure, white painted frame chair These Chairs are Stackable. The Load Capacity is 120 kg This chair is Pre Treated Epoxy Powder Coated in white colour
2	Doctor's Table	The size of the doctors table should be 1370mmWidthX 680mm Depth X 750mm Height with two storage units of 3 drawer unit and 1 unit of two drawers.It should be made from square tubular infrastructure with a large top surface. The product should have standard warranty with registered trademark
3	Duty Table for Nurses	Ambulatory table comes with 3 drawers which fits in hospital outpatient department Ambulatory Table is manufactured with no Sharp corners and is steel painted with epoxy powder coating. This table is robust construction,impact-resistant, scratch-resistant. It is easy to clean, washable and disinfectable. The overall dimension is 150 x 60 x 80 cm (L X W X H)
4	Table for Sterilisation use (medium)	
5	Long Benches (6 ½' x 1 ½')	Stainless steel long bench seat
6	Stool Wooden	
7	Stools Revolving	Frame is made of Round Tube The body section is mounted on 5 legs fitted with 5cms castors The height of the stool is adjustable with Pneumatic Gas Spring The Stool has Cushioned Seat Top
8	Steel Cup-board	
9	Wooden Cup Board	
10	Racks -Steel	90L x 37.5W x 195HFrame work made of Mild Steel Angle and Plate which are Pre Treated & Epoxy Powder Coated Open Rack is also known as Slotted Angle Rack It is used for Storage Purpose in Hospitals It comprises of Six Shelves
11	Patients Waiting Chairs	3 seater, Steel Size: 1900x630x900mm colour-blue Resistance against water Flat edges Highly very spacious (Airport three seater chair) colour -blue
12	Attendants Cots	Size is 1800mm in length x 600mm in width x 475mm in height with mattress Pretreated and epoxy coated material of construction Aluminium Composite Panel (ACP) sheet top Mounted on rubber shoes

Sl. No.	Name of the Equipment	specification
13	Office Chairs	Colour: Black Seat Size: 21'' W X 19'' D Back Size: 19'' W X 23'' H Overall Dimensions: 24'' W X 19'' D X 34'' - 36'' H Material: Fabric Pneumatic Seat Height Adjustment. Medium Height Back Rest. Lumbar Adjustment on Back Small Cushion Pillow Support on Back Nylon Netted Mesh Back Adjustable Up-Down Arm Rest. Virgin Quality PP Stand/Base Gas lift Height Adjustment. 90 Degree Tilt Lock. Symcro Push Back Design Machanism. Imported Castor Wheels with Good Grip. 360 Degree Revolving.
14	Office Table	Table Top made of 25mm thick particleboard with melamine finish All other parts made of 18 mm thick particleboard with PU finish Two tone colour combination as in image Lock on top drawer Holes for wire management on the top Product Dimensions & Weight Dimensions: mm (inch) 1500 (59) W x 750 (29.5) D x 750 (29.5) H PRODUCT DETAILS COLOR - Cherry/Black **brand-Nilkamal**
15	Foot stools	Frame made of Round Tube The body section is mounted on Rubber Shoes Foot Step is Double Step The top is fixed with Laminated non skid board Product Size: 45L x 45W x 45H cms
16	Filing Cabinets (for records)	Filing cabinet is made of MS Sheet It has 4 Drawers It is fitted with Lock Key
17	M.R.D. Requirements (record room use)	
18	Paediatric cots with railings	
19	Cradle	Frame Made of CRCA steel tube Cradle made of wire On PVC shoes Dimension: 900 L x 375 W x 975 H mm approx Pre-treated and Epoxy Powder Coated

Sl. No.	Name of the Equipment	specification
20	Fowler's cot	Specification: Overall size: 2100mm (L)*1050mm (W)*Ht Adj.590mm to 750mm (H) Features: CRCA tubular frame work Four section perforated CRCA sheet top Height adjustment , Backrest , Kneerest , Trendlengburg & reverse trendlengburg operated by screw and lever mechanism with attached S.S folding handles Stainless steel bows with coloured metal panel New design collapsible railings S. Telescopic saline rod with four locations Four heavy duty swivel castors of 125mm Dia.(two with break) Pretreated and epoxy powder coated
21	Ortho Facture Table	
22	Hospital Cots (ISI Model)	emi Fowler Bed Two section top. Removable high quality ABS head and foot penal. Backrest adjustable by smooth crank mechanism. Overall size : 2140 x 940 x 530mm Bed mounted on protective stumps. Collapsible safety stainless steel railing. Provision for I.V. Rod on all for location. Pretreated and Epoxy Coated Finish.
23	Hospital Cots Paediatric (ISI Model)	Complete with full-length dropside safety railings. Head and leg bows of equal height. Mounted on four 10cms. dia. castors (2 with brakes). Pre-treated and powder coated. Size : 137 x 76 x 60cm.
24	Wooden Blocks (Set)	
25	Back rest	Frame is made of Round Tube Back support is made of Canvas The Recline Angle is adjustable upto 4 positions with Rachet The Back rest is foldable Product Size : 50L x 40H cms
26	Dressing Trolley (SS)	Frame Work of Dressing trolley is made of Round tube It has stainless steel shelves with railing The body section is mounted on 7.5 cm castors The Dressing Trolley has push handle on one side It comes with stainless steel bowl size 32cm and Bottle Rack The Dressing Trolley is fitted with Drawer & Dust Bin Product Size:- 90 x 60 x 80cm
27	Medicine Almairah	Instrument Cabinet is made of MS Sheet It has 4 Glass Shelves The cabinet has 2 partitions The top partition has 2 Glass Door with Mild Steel frame The bottom partition has 2 doors made of Mild Steel It is fitted with Lock Key and Handle
28	Bin racks (wooden or steel)	

Sl. No.	Name of the Equipment	specification
29	ICCu Cots	**Electric ICU Bed-** Four section perforated two way longitudinal tilt hi-low facility 5 function electrically operated backrest, knee rest, hi-low, Trendlenburg and reverse Trendlenburg with battery backup Overall Size: 2140 x 940 x 500/760mm Removable high quality ABS head and foot panel with roller bumpers Collapsible safety stainless steel side railings 125mm Central locking castors two with brakes Provision for IV Rod on all four location Pre-treated and Epoxy coated
30	Bed Side Screen (SS-Godrej Model)	Frame Work of Bedside Screen is made round tubes It has 4 frames of equal size It is mounted on 6 castors of dia 5cm It comes with curtains and spring wire Product Size: 240L x 165H cm Knock Down Construction to save freight
31	Medicine Trolley (SS)	Medicine Trolley Cassette type is made of ABS Plastic with arc handle It has Double sides trays ,within 50 units medication casette bins It comes with one side plate cassette Medicine Trolley is easy to discharge It comes with noiseless castors , two with brakes and has centralized locking Multi-bin container is available at rear side for storage It has ergonomics designed push-handle for smooth movement Medicine trolley is equipped with total lock key system Product Size :78L x 49D x 92H cms
32	Case Sheet Holders with clip (S.S.)	Case Sheet Holder is made of Aluminium It is used to maintain the records of the patient
33	Bed Side Lockers (SS)	made of ABS Plastic It comes with one towel rail The body section is mounted on 5cm castors It comes with one drawer for storage Product Size : 45L x 45W x 75H cms
34	Examination Couch (SS)	Frame made of CRCA steel Cushioned Two section top Head section adjustable on Pneumatic pump Measuring 450 mm L x 510 mm W approx Body section admeasure 1380 L x 51 W mm approx Three drawers underneath the top section Three storage units under the draws Pull-out foot-step Foldable/adjustable B.P. Apparatus tray Dimension: 1870 L x 510 W x 810 H mm approx Pre-treated and Epoxy Powder Coated

Sl. No.	Name of the Equipment	specification
35	Instrument Trolley (SS)	Trolley is made of Mild Steel Sheet It has one drawer & 4 shelves It comes with oxygen cylinder cage & adjustable I.V. Rod The Monitor Trolley is mounted on 12.5cm castors It has a switch & socket panel for medical equipment Product Size :80L x 60D x 140H cms
36	Instrument Trolley Mayos (SS)	Frame Work of Instrument trolley is made of heavy angle The body section is mounted on 7.5 cms castors. It comes with Three Stainless Steel Shelves with upstands on four sides It comes with Push Handle on one side Knock Down Construction to save freight
37	Surgical Bin Assorted	
38	Wheel Chair (SS)	Ultra-lightweight T6 Aluminum S-Style Ergonomic Seat Frame Padded Seat and Back Upholstery Fixed Armrest with Padded Armrest Quick Release Axles and ForksHigh strength, light weight Swingaway Footrests
39	Stretcher/Patience Trolley (SS)	Frame work made of Tubular MS Pipe It has Safety Side Railing It comes with adjustable I. V. Rod, Utility Tray, Cylinder Cage The top is fitted with mattress The body section is mounted on 15 cms castors , two with brakes Size: 180L x 55W x 80H cms
40	Instrument Tray (SS) Assorted (10 Big Size)	
41	Kidney Tray (SS) - Assorted	
42	Basin Assorted (SS)	
43	(2 basin type)	
44	(1 basin type)	
45	o2 Cylinder Trolley (SS)	M.S. tubular pipe construction with two castors of 10cm. Pre-treated and powder coated.
46	Nitrous Cylinder Trolley (SS)	
47	Saline Stand (SS)	Stainless Steel I.V. stand with two locations & three wing base.Pretreated Epoxy powder coated I.V. stand with two or four locations. With Castors. Variable Height Adjustable.
48	Waste Bucket (SS)	
49	Dispensing table Wooden	
50	Bed Pan (SS)	
51	urinal Female	
52	Name Board for cubicals	
53	Waste Disposal - Bin/drums	Constructed from S.S. Removable inner container of plastic material. Capacity 11 lt. Size : 28 x 36cm.
54	Waste Disposal - Trolley (SS)	
55	Linen Almirah	
56	Stores Almirah	
57	Arm Board Adult	
58	Arm Board Child	

Sl. No.	Name of the Equipment	specification
59	SS Bucket with Lid	
60	Bucket Plastic	
61	Ambu bags adult with Mask Set	
62	o2 Cylinder with spanner and flowmeter ward type	
63	Diet trolley	stainless steel food warmer hospital food trolleys
64	Needle cutter and melter	
65	Thermometer clinical	Digital
66	Thermometer Rectal	Rectal Thermometer Digital, Prismatic Pear shaped bulb Range – 35-42°C & 95-108°F
67	Torch light	LED Type: 3 W White LED Product Size: 153*33*28mm Product Weight: 125g without battery and Battery weight will be 46 G And total package will be 450g Battery: 1x18650 Li-ion Battery Bulb Details: One 3W high power led in 120 lumens, white light. Three mode: High power, Low power, flashing. 16 steps of clear Zooming. Recharge Details: One 18650 Li-ion battery, 3.7V, 2200 mAh capacity Can be charged by AC charger or car charger (12V Optional) Charging time for AC charger:6 hours
68	Cheatles forceps assortted	
69	Stomach wash equipment	
70	Emergency Resuscitation Kit-Adult	
71	Laryngoscope adult and Child set	
72	Magill Forceps	
73	Airways 0-5 sets	
74	Refrigerator 165	LG 5 STAR WITH STABLIZER(2 YEAR WARENTY)
75	ILR	

Sl. No.	Name of the Equipment	specification
76	Deep Freezer	Equipped with CFC free refrigerator and air cooled compressor for providing efficient cooling as well as air flow in cabinet High capacity air cooled condensers that feature aerodynamically shaped fans Featuring washable concessive filter that provides for dust free usage System coming with thick polyurethane insulation of 50 - 80 mm for effective cooling support System require minimum power consumption System featuring presence of Digital electronic temperature controller cum indicator that provides for easy readability possibilities System construction featuring durable stainless steel interior that provides for maintaining temperature as well as allows easy cleaning Presence of double gasket seals between lid and cabinet that provides for increased efficiency of system Specifications: Working Range Available Upto - 20oC • Up to - 40oC • Up to - 60oC Electrical Requirement 220 – 240V/50Hz
77	Generator 125 kV	Fuel Type: Diesel Cylinders: 4 Phase: Three Phase Power Rating Range: 125 Fuel Tank Capacity Range: 200 Litres Weight Range: 2000 Speed: 1500 rpm Battery Capacity: 115 Ah Style: Standby Generators LxWxH(Weight): 3500x1200x1700 Rating (kVA): 125 Rated kW: 100 BHP: 152
	Grand Total	

Lab Equipment

S.No.	Item	specification
1	Bionocular Microscope	Viewing Head: Monocular/Binocular/Trinocular Head 45°, 360° rotatable Eyepiece: Anti Fungal Eyepiece WF10X Objective: Anti Fungal Achromatic Objective 4X, 10X, 40X Nosepiece: Quadruple Turret Analyzer: Rotatable Analyzer with Graduation 0°-90° Bertrand Lens: Bertrand Lens, sliding in/out of optical path Optical Compensator: Slip (first class red), ¼ slip, Quartz Wedge Revolving Round Stage: Diameter 135mm, 360° Rotatable & Graduated in 1° increments Minimum Resolution 6’ when using vernier scale center adjustable Focusing: Coaxial Coarse and fine Knobs Polarizer: Sliding in/out of optical path located on the top of collector Illumination: 6v20w Halogen Lamp
2	Electric calorimeter	Heating Coil: 5W Standard Binding Posts Coated Supporting Posts and Coil Single-Hole Rubber Stopper Calorimeter Connects to Power Supply, Ammeter, and Voltmeter

S.No.	Item	specification
3	electriccentrifuge,Table top	Max Speed : 14,000 rpm Max RCF : 20,290 (x g). Max Capacity : 6 x 50 ml (300 ml) CT14RD is our latest family member of CT series, having the following advanced features: Maintenance free brushless drive,power break Auto-lock rotor setting system Automatic rotor identification Powerful refrigeration system Large LCD display with informative graphics Wide selection of rotors
4	Hot plate	Overall size(W×D×H) mm 200×310×115Plate size(W×D) mm180×180(ceramic coated)Temperature (surface) 380???(Max.) Strring speed rpm 100 ~ 1,500 capacity m100 ~ 3,000 Heater(sheath) W680 Material top plate Ceramic coated top plate body Steel plate with powder coated finish Power source & consumption AC 230V 50/60Hz 1phase Size of Heating Pan: 10" x 12" Inches 10" x 16" Inches 12" x 18" Inches 18" x 24" Inches Temperature Range: up to 350 deg.C Energy Regulator: Yes Power Indicator: Yes Rating: 1500 to 4000W Power: 220 V AC 50/60Hz

S.No.	Item	specification
5	PH meter	pH Range: 0.00 to 14.00 Resolution: 0.01 pH Accuracy: ±0.01 pH Temp. Range: 0~100°C Resolution: 0.1°C Accuracy: ± 0.3°C pH Calibration Points: Up to 3 points with Auto-buffer Recognition mV Offset: ±150 mV Temp. Offset: ±5.0 °C pH Buffer Sets: USA (pH 4.01, 7.00, 10.01); NIST (pH 4.01, 6.86, 9.18) Temp. Compensation: Automatic/manual (0 to 100°C) Slope: % 80 to 120 pH Slope & Offset Display: Yes Hold Function: Yes Averaging / Stability (READY): Yes Memory: 50 Data Sets Inputs: BNC, ref. (half-cell), Ground, 2.5mm phono jack (ATC), power jack Output Recorder: (±2000 mV)
6	Padeatric Glucometer/Bilirubinometer	
7	glucometer	Automatic Glucometer Easy operation Compact designed Clear display

S.No.	Item	specification
8	Haemogllobinometer	Display -Digital 3½ Digit, Bright LED Display Mode% T, O.D. and 3 concentration Modes, C1,C2,C3Zero AdjustmentAutomatic (Without Knob rotation)Accuracy+ 0.02 O.D.Stability+ 0.02 O.D. Per HourLight Source6.8 Volt, 0.3 Amp Tungsten Lamp Amplifier Micro controller BasedMemoryIn built memory to retain last "K" factor for concentration even after power loss.Sample Volume1 ml. minimumPower 230 V + 10 % AC, 50 Hz.AccessoriesSet of Test Tube- 5 No.Dust Cover – 1 No.**Brand-Hellasi,TT3**
9	TCDC count apparatus	total number of counts 999.Counts in each group 0-99.No. Of operating keys. Twelve(12). BASO (Basophilic) Variate Lymphocyte EOS (Eosinophilic) MON (Monocyte) Neutrophilic stab granulocyte LYM (Lymphocyte) Prom(Promphocyte) Neutrophilic segmented Neutrophilic orthochromatic Normoblast Delete (--) or Minus (%) Percentage Reset Delete facility Available. Count category 9.Input voltage Mains -220 ±10%, 50Hz, Battery -12V DCPower Consumption 10W (Approx)
10	ESR stand withtubes	Special Pipe: stainless steel pipe Section Shape: Round (Westergen Method)
11	Test tube stand	Small, Normal, Large
12	Test tube rack	Small, Normal, Large
13	Test tube Holder	Normal
14	Spirit lamp	Normal (Bunsen Burner)/ Gas Nope

S.No.	Item	specification
15	Timer stop watch	Four countdown channels plus stopwatch channel and time-of-day clock. Timer/Stopwatch; Fisherbrand; Countdown; Traceable; Four-channel; Accuracy: 0.01%; Range: 100 hours; Display: 6-digit 0.75 in. H LCD display; Power supply: 2 year silver oxide; 2 oz. (57g); Count down, count, time out and time of day, extra-loud high decibel alarm Timer: Counts down from 99 hours, 59 minutes, 59 seconds to 1 second. Stopwatch: Counts up from 1 second to 99 hours, 59 minutes, 59 seconds. Timing capacity: 100 hours Accuracy: 0.01% “Thru-zero” function counts down, alarms, counts up Time-out mode Easel stand, lapel clip, hole for lanyard and magnetic back Alarm can be silenced automatically or manually Memory recall Time-of-day clock feature Dimensions: 6.35L x 6.03W x 1.27cmH (2.50 x 2.375 x 0.5 in.)
16	Alarm clock	Digital Lcd display battery and electrical
17	Ordinary Refrigerator	Capacity (Litres): 245 L Color: GREY Star Rating: 4 Star **Brand-LG**
18	knife shapner	
19	Whole blood finger Prick, HIV Rapid Test and	

List of Linen For MCH Wings

Sl. No.	Name of the Equipment	SPECIFICATION
1	Bedsheets	Color : blue/white/pink/green/yellow/purple/ peach gray with border(7 day colour coding bedsheet),Soft texture
2	Bedspreads	Material Pure Cotton Color -Blue
3	Blankets Red and blue	Pattern: Plain Size: 60" x 90" Blanket with thermal resistance of garments(TOG)
4	towels	Size 30 x 60 , 36 x 72 Color White Grey Material: Polyester / Viscose
5	Table cloth	COLOUR -RED
6	Draw sheet	Cotton Color Sky Blue Features Reusable,NO SHRINK
7	Doctor's overcoat	SIZE -,Small,Large,Free Size,Medium
8	Hospital worker OT coat	Navy blue
9	Patients house coat (for female)	Color: Green Size: Large, Small, Medium, Material: Pure Polyester, Mix of Polyester & Cotton Type: Stitched
11	Over shoes pairs	
12	Pillows	Shrink resistance Smooth finish Easy to clean
13	Pillows covers	Color: Blue Design: Plain Material: Cotton
14	Mattress (coir-foam) Adult	Specification of Rubberized coir mattress Material - Natural Latex and coir fibre. Brand - Ortho Medical
15	Paediatric Mattress	

16	Abdominal sheets for OT	Material Used: Non Woven · Weight: 38 GSM · Color: Violet · Special Feature: Water Repellent
17	Pereneal sheets for OT	Poly Material Sterility: Sterile Types: Plain Drapes
18	Leggings	
19	Curtain cloth windows and doors	designer courtain Long lasting shine * Perfect stitching colour-blue and gray mix printing
20	Uniform/Apron	
23	Mackin tosh sheet (in meters)	colour -red

A. Meternal Health Equipment for 100 Beded MCH Wing

S.No.	Item	specification
1	Labor Tabel with foot steps	CRCA steel tube frame SS three section top Leg section slides completely under the trunk section Trendelenburgs position by Pneumatic pump Adjustable back section on ratchet SS side railing, Hand grips on either side Padded Lithotomic crutches Mattress Dimension: 1830 L x 680 W x 760 H mm approx Pre-treated and Epoxy Powder Coated
2	Examination Table with foot steps	Frame made of CRCA steel Cushioned Two section top Head section adjustable on Pneumatic pump Measuring 450 mm L x 510 mm W approx Body section admeasure 1380 L x 51 W mm approx Three drawers underneath the top section Three storage units under the draws Pull-out foot-step Foldable/adjustable B.P. Apparatus tray Dimension: 1870 L x 510 W x 810 H mm approx Pre-treated and Epoxy Powder Coated
3	Pulse oxymeter adult	Low power consumption, battery-low indicator Adjustable brightness, Automatically power off Suitable for pediatric and adult Dual color LED displays SPO2, PR, Pulse bar and waveform Low power consumption
4	Oxygen Cylinder	
5	Sution Machine Adult-Electrical	Noiseless oil immersed pump Developed with ¼ Hp electric motor Easy mobility with smooth castors Self-sealing rubber caps on jars

S.No.	Item	specification
6	Foot Operated Suction Machine	Energy saving compressor Noiseless operation Sturdy stainless steel cabinet Mounted on heavy duty castors Specifications: Weight: 17kg. (Approx.) Dimensions: W xD xH = 12” x 9” x 36” (with Handle) Power Supply: 230V AC, 50 Hz.
7	Adult Resucitation ambu bag with r	Silicone Adult 1600 ml Resuscitator Set2 with 60CM H O Pop-off Valve• Silicone Mask Size: 4 No.• 2600 ml Oxygen Reservoir Bag• Disposable Oxygen Tubing 2 Mtr.• Reusable Latex Free
8	Newborn Rescusitation amu bag with mask 0,1 for NBCC	
9	Adult weighting Machine	Accurate reading Battery operated Easy usage
10	Digital newborn weighting machine	Power supply: AC 230V + 10%, -20%, 50 Hz, +-2% Overall dimensions: 555x385x240 mm with bowl Feather touch membrane keyboard
11	BP Apparatus	LCD display in any slanting position as per convenience Mark the blood pressure value when it is diagnosed by you with the help of stethoscope Automatic Deflation Valve No need to memorise your blood pressure reading LCD bar emulates the mercury column electronically Mercury free Accuracy similiar to Diamond mercurial instruments Memory for Systolic and Diastolic marking Pressure display on LCD column plus digital with background light Pulse rate readout Sturdy, durable, light weight, moulded cabinet Battery saving feature for LCD Superb control valve for perfect pressure drop Special cuff, bladder and bulb, more durable **Brand-Dimond** **LCD Bar BP Apparatus : BPDG 031, BPDG 234 LCD Super Delux**

S.No.	Item	specification
12	Room Warmer	Thermostat control for variable temperature settings Power(Watts): 1000W/2000W Voltage: 230V~50Hz **Brand-Bajaj Majesty RX 10 Fan Room Heater ,Usha Hc 812 T Thermo Fan Room Heater**
13	Foetal dopper	
14	Stethoscope	UltraScope Single StethoscopeWeight (oz) 5.80 Height 0.88 Length 30.00
15	Glucometer	Digital Automaticest Strips , Hospital Glucose Meter BRAND-**Accu-Check**
16	Wall Clock	Digital battery and electrical
17	Forceps Delivery Kit	
18	Silastic Vaccume Extractor	Silicone Cup : 40, 50, 60, 70 mm with/without release valve Stainless steel cup : 40, 50, 60, 70mm Traction handle, stainless steel / brass chrome plated Rubber / silicone tube, autoclavable Net powder / PVC coated to hold bottle Bottle glass moulded Rubber bung with 3 holdes. Stainless steel / brass chrome plated valve Stainless steel / brass chrome plated valve with pressure release valve Pressure guage Hand pump vacuum stainless steel / brass chrome plated with plastic / ms handle
19	Emergency Call Bell	wifi wireless bell
20	Delivery Tray	Containing Scissor 1 , Artery Forceps 2, Sponge holding forceps 2, Sims Speculum 1, Bowl for antiseptic lotion 1, Kidney Tray 1
21	Episiotomy tray	Episiotomy Scissor 1 , Artery Forceps 2, Allis Forceps 2, Sponge holding forceps 1, Toothed Forceps 1, Thumb forceps 1, Kidney Tray 1 , needle holder 1, needle (round body 1 and cutting 1)

S.No.	Item	specification
22	LSCS Set	Instrument Tray with cover – Stainless steel 300x 200 x 65 -1 Towel Clips (Backhaus) 13cm- 6 Sponge Holding Forceps (Forester; Straight; serrated) 25cm -2 Artery Forceps Straight – CRILE (Small) 14 cm- 6 Halstead Mosquito Forceps 12.5 cm -6 Tissue Forceps 18 - 20 cm- 4 Needle Holder (Mayo-Hegar) Straight 18 cm -2 Surgical Knife Handle (one each of No. 3 and 4 to be used with surgical blades sizes 20,21,22,23 and No 4 to be used with blade sizes 10,11,12 and 15-2 sets Abdominal Retractor size 3 -1 set Abdominal Retractor Double Ended (Richardson-Eastman) set of two stainless steel blades – Large - 49mm wide x 63mm deep & 38mm wide x 49mm deep, total length – 28cm Small – 20mm wide x 28mm deep & 28mm wide x 36mm deep, total length – 26 cm-1 Abdominal Retractor Double Ended (Richardson-Eastman) 1set of two stainless steel blades – Large - 49mm wide x 63mm deep & 38mm wide x 49mm deep, total length –28cm Small – 20mm wide x 28mm deep & 28mm wide x 36mm deep, total length – 26cm Bladder Retractor "KOCHER" 80x65mm wide, 25cm length -1
23	T.T. Set	
24	D & C Set	
25	MVA Syringe	
26	Kocker Forceps small	
27	kocker Forceps Large	
28	Sims Speculam Large	Brass & Stainless (1 inch 2 inch 3 inch)
29	Vaginal Wall Retractor	Sizes 10" Material:Stainless Steel • Double Ended, Serrated • Rust free • Sturdy construction • Accurate dimensions

S.No.	Item	specification
30	Uterine Sound	Stainless Steel Length 12-1/2" Tip Configuration Blunt Instrument Type Uterine Sound Curvature Flex Working Surface Style Smooth Handle Thumb Grip Material Silver Plated Disposable or Reusable Reusable Sterile or Non-Sterile Non-Sterile
31	PPIUCD Kit	
32	Cervical Biopsy Set	INSTRUMENTS BOX STAINLESS STEEL 32X15X6cm / 12½x6x2¼" MUSEUX TENACULUM FORCEPS 24.0cm / 9 ½", 8mm FAURE CERVICAL BIOPSY AND SPECIMEN FORCEPS 21.0cm / 8 ¼" SCHUBERT CERVICAL BIOPSY AND SPECIMEN FORCEPS 29.0cm / 11 ½" RECAIMER-SIMS-BUMM UTERINE CURETTE 31.0cm / 12 ½", MALLEABLE, FIG. 00-16 COLLIN VAGINAL SPECULA, 110 x 40mm FOERSTER SWAB FORCEPS 25.0cm / 10", STR, SERRATED SIMS UTERINE SOUND 32.0cm / 12 ½", MALLEABLE
33	Formaline Dispenser	
34	Air Purifier for LR, OT, SNCU	Dual UV sterilization air purifierDual UV sterilization air purifier Purification process: 6- stage purification process, electric static / hepa filter / UV, photo catalyst / activated Carbon / ions Negative ions: Produces 8,000,000 ions/cm3 Noise levels: 45 db Voltage supply: 220 V, 50 Hz Area covered: 500 Sq. ft. approx **or** 250 sq. ft Disk surface: 1.7 sq. yd.Air throughput: up to 2 cu. yd/min.Speeds: 2Sound level: 37/42 dBA Dimensions: 9.5 x 10.5 x 11.8 in Weight: approx. 7 lb Motor ouput: approx. 32 watts Tank capacity: 1.2 gallons Request Callback
35	Suturing Set	Forceps, Artery, Curved, 150mm Forceps, artery , Straight, 150mm Forceps, Dissecting, Straight, Plain, 130mm Forceps, dissecting, Straight, 1/2 Teeth, 130mm Forceps, Sponge Holding, 240mm Needle Holder, mayo, 150mm Scissors, mayo, Straight, 150mm Scissors, ligature, Spencer, 115mm Instrument Container, s/s with cover
36	L.P.Tray Set	

S.No.	Item	specification
37	Endo Tracheal tubes Sets	Advanced Micro-Thin Polyurethane Cuff Material Designed to Reduce Micro-Aspiration Endotracheal tube - cuffed: Size: 4, 4. 5, 5, 5. 5, 6, 6. 5, 7, 7. 5, 8, 8. 5,9, 9. 5,10(mm) Endotracheal tube - plain: Size: 2, 2. 5, 3, 3. 5, 4, 4. 5, 5, 5. 5(mm)
38	Tongue depressors	6 inch long ,wood